



Test date 17/SET/2010 : Vacuum pump + Workholder + Heat controller

- ✓ The time to reach the nominal temperature (170 ° C for this application) is 18 minutes.
- ✓ Workholder power is 180W.
- ✓ The system vacuum is OK